

Test Certificate

concerning the test of influences on MRI systems caused by non-Siemens products

1. Test certificate, product designation and manufacturer

This is to certify that the following product

IMROC-Interventional MR Optical Communication System
tested with IMROC EOU 6000 Interface, serial number 16-E-2501,
IMROC Mixing Console, serial number 16-C-2801,
OptoACTIVE slim headphone and FOMRI microphone.

Of the manufacturer:

Optoacoustics Ltd
17 Hata'asia Street
60212 Or-Yehuda
Israel

does not impair the functioning of the MRI systems of Siemens AG, Healthcare Sector, Magnetic Resonance Systems, specified in section 2:

Potential functional restrictions of the MRI systems are indicated in section 4.

Possible adverse effects of the MRI systems on the product indicated above are expressly not subject of the test on which this certificate is based. Therefore, this certificate does not imply non-interference of the above-mentioned product through the MRI system.

2. MAGNETOM systems involved

Seq. no.	MR System or option (type designation)
1	Skyra (3 Tesla)
2	Verio (3 Tesla)
3	Trio a Tim System (3 Tesla)
4	Trio (3 Tesla)
5	Biograph mMR (3 Tesla)
6	Aera (1.5 Tesla)
7	Espreo (1.5 Tesla)
8	Avanto (1.5 Tesla)

- 9 Symphony a Tim System (1.5 Tesla)
- 10 Symphony (1.5 Tesla)
- 11 ESSENZA (1.5 Tesla)

3. Applications

The IMROC System is intended for use for the following purpose in connection with magnetic resonance application; the test certificate refers to this purpose:

- Interventional MR Optical Communication System

4. Restrictions

During the use of this product the following restrictions of functionality and/or applications of the MRI systems, options and accessories apply:

- n.a..

5. Warnings

The following safety precautions must be observed when using this product in connection with the MRI systems listed in section 2:

- Use the IMROC EOU 6000 Interface and IMROC Mixing Console only outside the MRI Magnet room. A warning note must be included in the operation manual and affixed on the equipment.

6. Validity

This test certificate shall be valid until revoked by Siemens AG, Healthcare Sector.

Erlangen, 2011-04-15

Siemens AG



Weichselbaum



Brandl